



nutrilab

Certificate of analysis

Order-no : 875491

Kiwa Nederland B.V.
Attn. Mrs. C. Geenen
Wilmsdorf 50
7327 AC APELDOORN

Description	:	MEDIZINISCHER MUND & NASENSCHUTZ	
Client project-no	:	200901945	
Sample received	:	02-10-2020	Sample date : 01-10-2020
Report date	:	15-10-2020	Start date microbiology : 06-10-2020
Sampling	:		Sample delivery : Nutrilab BV
Packing	:	Original	Sample temperature : Room temperature
Sealed	:	No	Sample condition : Sample and packing intact
Remark	:	Order number : CG-201001-02	

Test	Result
Bacterial Filtration Efficiency (BFE)	
I 35000 Testing BFE (n=5) (equiv. NEN-EN 14683+C1)	
I 35050 Test conditions	
I 35051 Dimensions of test specimens (width x height)	17,5 x 9,5 cm
I 35052 Size of the area tested (width x height)	48.0 cm ²
I 35053 Side facing the aerosol	Face side
I 35054 Flow rate during testing	28.3 ± 0.3 L/min
I 35060 Mean of the total plate counts of the two positive controls	1,972 cfu
I 35070 Mean plate count of the negative controls	0 cfu
I 35100 Bacterial Filtration Efficiency (BFE, equiv. NEN-EN 14683+C1)	
I 35101 BFE specimen 1	> 99.7 %
I 35102 BFE specimen 2	> 99.7 %
I 35103 BFE specimen 3	> 99.7 %
I 35104 BFE specimen 4	> 99.7 %
I 35105 BFE specimen 5	> 99.7 %
I 35199 Average BFE	> 99.7 %
Splash resistance	
I 35600 Splash resistance (equiv. ISO 22609:2004)	
I 35611 Splash resistance specimen 1	Pass 16 kPa
I 35612 Splash resistance specimen 2	Pass 16 kPa
I 35613 Splash resistance specimen 3	Pass 16 kPa
I 35614 Splash resistance specimen 4	Pass 16 kPa
I 35615 Splash resistance specimen 5	Pass 16 kPa
I 35616 Splash resistance specimen 6	Pass 16 kPa
I 35617 Splash resistance specimen 7	Pass 16 kPa

Page 1 / 3

Nutrilab bv
Postbus 7, 4284 ZG Rijswijk
Burgstraat 12, 4283 GG Giessen
t. (0183) 44 63 05
f. (0183) 44 25 97

info@nutrilab.nl
www.nutrilab.nl
KvK 18114291
BTW NL002007654B01

Explanation of symbols:

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub-contractor





I 35618	Splash resistance specimen 8	Pass	16 kPa
I 35619	Splash resistance specimen 9	Pass	16 kPa
I 35620	Splash resistance specimen 10	Pass	16 kPa
I 35621	Splash resistance specimen 11	Pass	16 kPa
I 35622	Splash resistance specimen 12	Pass	16 kPa
I 35623	Splash resistance specimen 13	Pass	16 kPa
I 35624	Splash resistance specimen 14	Pass	16 kPa
I 35625	Splash resistance specimen 15	Pass	16 kPa
I 35626	Splash resistance specimen 16	Pass	16 kPa
I 35627	Splash resistance specimen 17	Pass	16 kPa
I 35628	Splash resistance specimen 18	Pass	16 kPa
I 35629	Splash resistance specimen 19	Pass	16 kPa
I 35630	Splash resistance specimen 20	Pass	16 kPa
I 35631	Splash resistance specimen 21	Pass	16 kPa
I 35632	Splash resistance specimen 22	Fail	16 kPa
I 35633	Splash resistance specimen 23	Pass	16 kPa
I 35634	Splash resistance specimen 24	Pass	16 kPa
I 35635	Splash resistance specimen 25	Pass	16 kPa
I 35636	Splash resistance specimen 26	Pass	16 kPa
I 35637	Splash resistance specimen 27	Pass	16 kPa
I 35638	Splash resistance specimen 28	Pass	16 kPa
I 35639	Splash resistance specimen 29	Pass	16 kPa
I 35640	Splash resistance specimen 30	Pass	16 kPa
I 35641	Splash resistance specimen 31	Pass	16 kPa
I 35642	Splash resistance specimen 32	Pass	16 kPa

Microbial cleanliness (Bioburden)

I 35200	Microbial cleanliness (Bioburden) (equiv. NEN-EN 14683+C1)			
I 35210	Total bioburden per individual mask			
I 35211	Total bioburden specimen 1	Q	< 30	cfu/mask
I 35212	Total bioburden specimen 2	Q	< 12	cfu/mask
I 35213	Total bioburden specimen 3	Q	< 12	cfu/mask
I 35214	Total bioburden specimen 4	Q	< 12	cfu/mask
I 35215	Total bioburden specimen 5	Q	< 30	cfu/mask
I 35250	Total bioburden per gram			
I 35251	Total bioburden specimen 1	Q	< 8	cfu/g
I 35252	Total bioburden specimen 2	Q	< 3	cfu/g
I 35253	Total bioburden specimen 3	Q	< 3	cfu/g
I 35254	Total bioburden specimen 4	Q	< 3	cfu/g
I 35255	Total bioburden specimen 5	Q	< 8	cfu/g

Conclusion*

I 35290	Mask type based on BFE performance requirements for medical face masks is:	I/II/IIR
---------	--	----------





nutrilab

Certificate of analysis

Order-no : 875491

I 35690	Mask type based on splash resistance performance requirements for medical face masks is:	I/II/IIR
I 35280	Mask type based on bioburden performance requirements for medical face mask is:	I/II/IIR

*) Performance requirements for medical face masks (acc. European Standard no. EN 14683:2019+AC):

Bacterial filtration efficiency (BFE) (%): Type I: ≥ 95 , Type II: ≥ 98 , Type IIR: ≥ 98
Differential pressure (Pa/cm²): Type I: < 40 , Type II: < 40 , Type IIR: < 60
Splash resistance pressure (kPa): Type I: n.a., Type II: n.a., Type IIR: ≥ 16.0
Microbial cleanliness (cfu/g): Type I: ≤ 30 , Type II: ≤ 30 , Type IIR: ≤ 30

Remark:

Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

This certificate of analysis is a test report. The tested samples are part of the mentioned batch/lot number. Batch validation is not the scope of this report.

Start date analysis: 06-10-2020, end date: 15-10-2020.

Findings are based on the sample as submitted. For more detailed information on an applied method and the corresponding measurement uncertainty please contact our Customer Service.

Any interpretation of analytical results mentioned on this certificate lies outside the scope of accreditation.

With the unit % is meant: w/w%, unless otherwise stated.

Nutrilab is not responsible for the information provided by the client.

Only reproduction of the entire certificate is permitted without explicit written authorisation by Nutrilab BV.

Operational manager : ing. T. Boogaard



Page 3 / 3

Explanation of symbols:

Q RvA accredited test (ISO / IEC 17025)
I Test performed by Nutrilab BV
E Test performed by sub-contractor



Nutrilab bv

Postbus 7, 4284 ZG Rijswijk

Burgstraat 12, 4283 GG Giessen

t. (0183) 44 63 05

f. (0183) 44 25 97

info@nutrilab.nl

www.nutrilab.nl

KvK 18114291

BTW NL002007654B01